



財團法人全國認證基金會
Taiwan Accreditation Foundation

Certification Accreditation

(Certificate No : L1328-230119)

This is to certify that

Taiho Tool Mfg. Co., Ltd.
Calibration Laboratory

NO.23, Bengong 1st Rd., Gangshan Dist., Kaohsiung City, Taiwan (R.O.C)

is accredited in respect of laboratory

- Accreditation Criteria** : ISO/IEC 17025:2017 ; CNS 17025:2018
Accreditation Number : 1328
Originally Accredited : March 01, 2005
Effective Period : March 01, 2023 to February 28, 2026
Accredited Scope : Calibration Field, see described in the Appendix



Scan to verify

Ching-Chang Lien

Ching-Chang Lien
President, Taiwan Accreditation Foundation
January 19, 2023

Accreditation Number : 1328

Laboratory Head : CHEN, Fu-Chun

Length

calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KA1002 plug gauge	OSG/Master Plug Gauges /Class 1	plug gauge diameter calibration method (Document No.: QET03)	1	mm	50	mm	major diameter	1.1	μm
Approval Signatory: WU, Jui-Fang; KANG, Chia-Sheng; CHEN, Fu-Chun									
KA1008 screw thread plug gauge	Master Plug Gauges/OSG /Class 1 Three Wires /EISEN/TW	screw thread plug gauge major diameter calibration method (Document No.: QET01) screw thread plug gauge pitch diameter calibration method (Document No.: QET02)	M1		M50		major diameter	1.1	μm
			M1		M50		pitch diameter (thread angle: 55°)	4.0	μm
			M1		M50		pitch diameter (thread angle: 60°)	3.8	μm
Approval Signatory: WU, Jui-Fang; KANG, Chia-Sheng ; CHEN, Fu-Chun									

Note: Smallest uncertainty represents an expanded uncertainty using a coverage factor approximately 95 % level of confidence.
(Null Below)

